Tel.: +44 (0) 1525 218146 mipuk.co.uk



The Softer Side of Healthcare

UK EC DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer and is represented by the undersigned.

MIP-UK Ltd, 2nd Floor, Clipper House, Leighton Industrial Park Billington Road, Leighton Buzzard, Bedfordshire, LU7 4AJ, England

hereby declares that the product

Type of product(s): Patient Transfer and Slide sheets

Basic UDI-DI: **7540289R175371639J**

GMDN Code: **37163**

Intended Purpose: A manual device used by healthcare professionals to assist in the physical

movement and transfer of a person/patient (e.g., a hospitalized, a person with a disability, or a geriatric) from one position to another. The device typically has no lifting capabilities and utilizes sliding techniques to move the person. It is typically designed using low-friction artificial fibres and may be designed as an open roll, rather like "a sleeping bag", or a flat sheet. It is slid under the person to be moved, the person is gently pushed by the attending

staff and it rolls whilst the person glides on the rotating surface.

Model(s):

CS366-0001	CS413-001	CS441-0001	CS567-0001
CS990-B001	JC100-001	JC100-0011	LP003-0001
LP004-0001	LP004-0004	LP004-0006	LP004-0007
LP004-0009	LP004-0011	LP004-0012	LP004-RY07
LP010-0001	LP010-0002	LP010-0003	LP010-0004
LP010-0005	LP010-0006	LP010-0007	LP010-0008
LP010-0009	LP010-0010	LP011-0001	LP011-0002

Tel.: +44 (0) 1525 218146 **mipuk.co.uk**



The Softer Side of Healthcare

LP011-0003	LP011-0004	LP011-0005	LP011-0006
LP011-0007	LP011-0008	LP011-0009	LP011-0010
LP012-0006	LP012-0009	LP012-0016	LP012-0106
LP012-0109	LP012-0116	LP012-0207	LP012-P009
LP020-0003	LP020-0011	LP021-0003	LP021-0005
LP025-0001	PTD-56/BA/LG/UK	PTD-56/BA/Y/UK	PTD-LS/S/BA1/Y/UK
PTD-LS/S/UK	PGF20085	PGF200100	PGF100100
PGF150100	PGF220140	PGT150100	PGT100100
PGT200100	PGT220140	PSF20085	PSF200100
PSF200140	PSF100100	PSF150100	PSFH20085
PSFH200100	PSFH200140	PSFH100100	PSFH150100
PST8585	PST11585	PST130100	PST14585
PST20085	PST200100	PST150100	PST220140
RSF20085	RSF200100	RSF200140	RSF100100
RSF150100	RSFH20085	RSFH200100	RSFH200140
RSFH100100	RSFH150100	RST8585	RST11585
RST130100	RST14585	RST20085	RST200100
RST220140	PSTL20070	PSTL200140	PSTL150120
RSTL200140	CS555-0001		

was designed and manufactured in conformity with the following legislation(s)

- UK Medical Devices Regulations 2002
- REACH Regulation (EC) No. 1907/2006

2nd Floor Clipper House Leighton Industrial Park Billington Road Leighton Buzzard, Bedfordshire LU7 4AJ England

Tel.: +44 (0) 1525 218146 **mipuk.co.uk**



The Softer Side of Healthcare

The product(s) has been classified according to Annex IX of Directive 93/42/EEC, read with Directive 2003/12 and Directive 2005/50 as:

Class I Medical Device (non-invasive devices Rule 1)

Applicable standard(s):

- EN 15223-1:2016: Medical devices Symbols to be used with medical device labels, labelling and information to be supplied
- BS EN 14971:2019: Medical devices. Application of risk management to medical devices

Certifications:

ISO 9001:2015 Certificate # 40662020 Issued by: QMS International ISO 14001:2015 Certificate # 14130892 Issued by: QMS International

Leighton Buzzard (2021-08-09)

Paul Crabtree Managing Director / MIP UK