freeway

T60 Shower Chair



User Manual

1 Introduction 4
1.1 Manufacture 4
1.2 Symbols Used 5
1.3 Contraindications / Limitations 6
1.4 Intended Use 6
1.5 Safety Instructions and Warnings 7
2 Components/Key Parts 8
2.1 Chair Configurations 9
3 Accessory List 10
4 Assembly Instructions 12
4.1 Attaching the seat 12
4.2 Attaching the backrest 12
4.3 Attaching the armrest 13
4.4 Attaching the footrest 13
4.5 Attaching accessories 13
4.5.1 Buckle or VELCRO ® Brand fastener Waist Strap 13
4.5.2 Full Adult Safety Harness 14
4.5.3 Toe Straps with Velcro ® Brand fastener 14
4.5.4 Heel Strap/ Heel and Toe Strap combine 15
4.5.5 Ankle Supports 15
4.5.6 Calf and Heel Support 15
4.5.7 Dartex Footrest Covers 16
4.5.8 Wrap around Dartex Footrest Covers 16
4.5.9 Pelvic Strap 16
4.5.10 Vinyl Calf Support 17
4.5.11 Clip on Padded Armrests 17
4.5.12 Clip-on Padded Foot Trays /Individual Clip-on Padded Trays 17
4.5.13 Detachable Headrests 18
4.5.14 Curved Padded Head Support 18
4.5.15 Head Support Cushion 18
4.5.16 Thoracic Back Pad with Strap and Wedges 19
5 Operating Instructions 19
5.1 Operating the Braked levers 19
5.2 Footrest adjustment 20
5.3 Using the Commode Pan (Accessory) 20
5.4 How to use the Harnesses 20
5.5 How to use and adjust the waist straps 21
5.6 How to use and adjust the pelvic straps 21
5.7 How to use and adjust the Ankle Supports 21
6 Technical Specification 22
6.1 Shower Chair Frame Dimensions 22
6.2 Shower Chair Aperture Dimensions 23
7 Environmental Conditions 24

7.1 Operating Environment 24
7.1.1 Normal Operating Conditions 24
7.2 Storage Conditions 24
7.2.1 Shipping and Storage Conditions 24
8 Daily Checks 25
9 Cleaning 26
9.1 Metal Frame Cleaning 26
9.2 Fabric Cleaning 26
9.3 Disinfecting (if necessary) 26
10 Servicing 28
11 Troubleshooting 29
12 Disposal 29
13 Warranty 30
14 Service Record History 31

1 Introduction



This user manual should be kept safe for future reference.



Please read and understand this manual in its entirety before using your Freeway Shower Chair.

The information in this manual is important for the safety of anyone near the shower chair and must be read and understood to help prevent injuries. It is also crucial to the proper operation and maintenance of the shower chair.

Should any questions arise from reviewing this manual, contact your local authorised representative.

Failure to comply with warnings in this manual may result in; injury to the operator and/or client and/or damage to the shower chair or related components.

If, during the use of this device or because of its use, a serious incident has occurred please report it to the manufacturer and to your national authority.

Store this manual with the documents included with the shower chair. Contents of this manual are subject to change without prior written notice.



Do not attempt to use this equipment without first understanding the contents of this manual.



Unauthorised modifications on this product may affect its safety. The manufacturer will not be held responsible for any accident, incident or deficiencies of performance that occur because of any unauthorised modification to its products.

1.1 Manufacture

The shower chair is manufactured at the address below:



Prism Medical UK
Unit 1, Tir Llwyd Industrial Estate, St Asaph Avenue, Kinmel Bay, Conwy, LL18 5JZ
Telephone number: 01924 840100.

1.2 Symbols Used

The Table below includes all symbols from BS EN ISO 15223-1:2021 that can be found in this manual and on the Product and what they represent. Refer to this Table when you are unsure of what a symbol represents.



 Place of Manufacture	 Country of Manufacture	 Serial number	 Catalogue Number
 Caution – see instructions for use	 Consult instructions for use	 UK Conformity Assessed	 Safe Working Load
 For internal use only	 Packaging indicator – Keep dry	 Do not use if package is damaged	 Packaging indicator – This way up
 Fragile, handle with care	 Please observe local laws on recycling	 Non-sterile	 Atmospheric pressure limitation
 Temperature range	 Humidity range		

Table 1-1

1.3 Contraindications / Limitations

There are no known "contraindications" associated with the usage of the shower chair, provided it is used as per the manufacturer's recommendations and guidelines. However, it is recommended that a client specific assessment is completed by a trained and knowledgeable health care professional to determine the method of transfer and use.

The manufacturer does not recommend a required number of operators for the use of our products. This information and recommendation can only be provided after a thorough personalised, case specific assessment, as there are many factors that can influence these decisions.

Limitations of the chair include:

- The chair should only be operated by individuals who are competent and have received appropriate training on its use.
- The chair should only be used with users weighing under the safe working load of the chair.
- The chair is only to be used within the environment it is intended for. Chairs must only be relocated by an authorised person.
- The chairs are designed for human transfer/use only. There is no other application to this product.
- If the chair is attendant propelled, the operator of the chair must always pay attention to the well-being of the user. Users should not be left unsupervised during operation.

1.4 Intended Use



For internal use only.

The shower chair, with a SWL (safe working load) of 200kg is designed to provide safe, and comfortable seating and showering of a user. The chair is an easily manoeuvrable with easy-to-use hygiene solution.

The shower chair's primary use is to provide the user with an effective solution for their toileting and showering needs. The chair can be used to transfer the user into the showering room where they can either transfer to a shower stretcher for bathing or receive a shower while remaining in the chair. The chair is also designed to locate over a toilet with an aperture seat to allow the user to use the toilet from the chair. It's important to note that the chair is designed as standard for use over a toilet of approximately 420mm from the floor to the top of the toilet. The chair provides adaptable options, including different style apertures, to ensure maximum user comfort during use.

The manufacturer always recommends training before use of our products. For more details on the training that can be provided please contact your local authorised dealer.



The device is used under instruction, and the operation of the aid is undertaken by a trained carer. A risk assessment must be performed before using the chair to ensure safe use can be established.



You may need to seek specialist advice on how to assist some people with specific moving and handling needs. Sources of advice include, but are not limited to, professional bodies and organisations, occupational therapists, physiotherapists, manual handling advisers and ergonomists with experience in health and social care.

1.5 Safety Instructions and Warnings



Ensure to read and understand all the statements below, for the safety of the caregivers and users, along with warranty requirements.

- If you are unsure on the correct use of this product, please contact the manufacturer or a professional for further information or training.
- The shower chair must only be operated (placing the user into/out of the chair) on a flat and levelled surface.
- The shower chair should be used to transfer a person on a level surface, so do not exceed a slope of 5˚.
- Exercise caution when self-propelling the chair, and avoid moving at high speed, as this can increase the risk of an accident.
- All shower chairs have less grip when pushed on a wet surface; be vigilant of the surroundings.
- Under no circumstance should the shower chair be put in control of a person who has not been properly trained in the use and care of this equipment. Failure to adhere to this warning may result in serious injury to the operator and/or the user.
- The shower chair and associated accessories are not toys. Do not use it for unsafe practices. Do not allow children to play with the shower chair or any of its components. The shower chair should not be used for any practice except its intended use.
- In facilities where more than one operator will be responsible for using the shower chair, it is important that all such members be trained on the shower chair prior to use. A training program should be established by the facility to acquaint new operators with this equipment.
- Your guarantee is void if persons unauthorised by the manufacturer perform work on the shower chair.
- To maintain optimum function, the shower chair should be inspected and maintained on a regular basis. See section 'Daily checks, Servicing and Cleaning' within this user manual.
- A list of approved accessories is listed in this user manual.
- The shower chair and the associated accessories are intended only for showering and transferring of a person. The manufacturer will not be responsible for any damage caused by the misuse, neglect, or purposeful destruction of the equipment and/or its associated components.
- Any accessories used with the shower chair should be checked before each use to ensure that they are in good working order. Check for signs of wear and ensure that all labels are legible. Report any unusual wear to your local authorised dealer.
- The shower chair and its associated accessories are certified to a maximum load of 200 kg. Do not exceed the maximum rated load of any of the components.
- Do not stand on the footrests; they are not designed to bear the full weight of the user.
- Ensure that a clear space is maintained around the shower chair. Before using the shower chair always check for and move away any obstacles.
- The shower chair is not intended to be used as a seating product for extended periods of time.
- When the shower chair is not in use, apply the brakes to prevent unwanted lateral movement.

2 Components/Key Parts

Please see below to familiarise yourself with the components of the shower chair. If you have not received all the components listed below, contact your local dealer immediately – contact details are provided on the last page of this manual.

Item	Description
1.	Frame
2.	Backrest
3.	Armrests
4.	Seat
5.	Footrests
6.	Self-Propelling Wheel
7.	Castors
8.	Serial Number Label (Located on frame under seat)
9.	Safe Working Load Label
10.	User Manual

Table 2-1




Fig. 2-1

2.1 Chair Configurations

The shower chair is available in various configurations, including width, frame type, seat depth, seat type, armrest type, backrest type and footrest type. See list below for the standard options.

Frame Options				
Frame Type	Frame Width (A-Large) (B-Medium) (C-Small)			Benefits
Standard Frame	A	B	C	Standard chair to be used with standard 16"/440mm height toilets

Table 2.1-1

Seat Options				
Seat Type	Seat Width (A-Large) (B-Medium) (C-Small)			Benefits
Fully Padded	A	B	C	Standard seat providing maximum comfort
Horseshoe	A	B	C	Horseshoe shaped aperture providing improved access
Full Ring	A	B	C	Provides aperture while maintaining full seat comfort
Vinyl Covering	√	√	√	Durable
Dartex Covering	√	√	√	Softer material

Table 2.1-2

Backrest Options	
Backrest Type	Benefits
Standard Push Handle Backrest	Personal preference
Standard Vertical Handle Backrest	Personal preference
5˚ Reclined Backrest	Provides a 5˚ recline
Extended Backrest	For users who require extended backrest for support

Table 2.1-3

Armrest Options	
Armrest Type	Benefits
Standard Armrest	Standard swing away armrests
Half Wide Armrest	Provides an additional 50mm width
Wide Armrest	Provides an additional 100mm width

Table 2.1-4

Footrest Options	
Footrest Type	Benefits
Standard Footrest	Standard folding footrest
Heavy Duty Footrest	Sturdier aluminium footrest

Table 2.1-5

Additional configurations are available upon request. Please contact your local Dealer for more information.

3 Accessory List

The list below shows a list of approved compatible accessories that can be ordered at any time during the life of the shower chair. Contact your local dealer to order these parts.

Straps and Harnesses		
Description	**Part number**	**Comments**
540mm Waist band VELCRO ® Brand fastener	880409	For size A width frame
490mm Waist band VELCRO ® Brand fastener	880412	For size B width frame
440mm Waist band VELCRO ® Brand fastener	880418	For size C width frame
540mm Waist band buckle fastening	880408	For size A width frame
490mm Waist band buckle fastening	880411	For size B width frame
440mm Waist band buckle fastening	880417	For size C width frame
Comfort Pad for 540mm Waist Strap	880413	For A size width frame to fit behind buckle/ VELCRO ® Brand fastener waist strap
Comfort Pad for 490mm Waist Strap	880414	For B size width frame to fit behind buckle/ VELCRO ® Brand fastener waist strap
Comfort Pad for 440mm Waist Strap	880415	For C size width frame to fit behind buckle/ VELCRO ® Brand fastener waist strap
540mm - Pelvic Strap & Comfort Pad	880407	For size A width frame
490mm - Pelvic Strap & Comfort Pad	880410	For size B width frame
490mm - Pelvic Strap & Comfort Pad	880416	For size C width frame
Full Harness	880423	For Postural support
Heel/Calf Strap	880419	Assist users' foot on footplate
Heel Support	910163	Assist users' foot on footplate
Heel supports and Toe Strap Combine	910195	Assist users' foot on footplate
Toe Straps	880400	Assist users' foot on footplate
Ankle Support Large	880402	Assist users' foot on footplate
Ankle Support Medium	880403	Assist users' foot on footplate
Ankle Support Small	880405	Assist users' foot on footplate
Paediatric Ankle Support	880404	Assist users' foot on footplate
Backrest Accessories		
Description	**Part number**	**Comments**
540mm Detachable Headrest	910099	For size A width frame (Only available with extended height backrest)
490mm Detachable Headrest	910100	For size B width frame (Only available with extended height backrest)
440mm Detachable Headrest	910101	For size C width frame (Only available with extended height backrest)
540mm Detachable Headrest to take Curved Head Support	910138	For size A width frame (Only available with extended height backrest)
490mm Detachable Headrest to take Curved Head Support	910137	For size B width frame (Only available with extended height backrest)
440mm Detachable Headrest to take Curved Head support	910136	For size C width frame (Only available with extended height backrest)

Flow Head Support Cushion	155805	For size A, B & C width frames (Only available with extended height backrest)
Curved padded head support	910185	Only available with extended height backrest and detachable headrest to take head support
25mm Padding for Headrest	910485	For Size A width detachable headrest to take padding
50mm Padding for Headrest	910486	For Size A width detachable headrest to take padding
25mm Padding for Headrest	910487	For Size B width detachable headrest to take padding
50mm Padding for Headrest	910488	For Size B width detachable headrest to take padding
25mm Padding for Headrest	910489	For Size C width detachable headrest to take padding
50mm Padding for Headrest	910490	For Size C width detachable headrest to take padding
25mm Thoracic Backrest padding with Straps	770195	Retrofit to Size A standard backrest for wedges
25mm Thoracic Backrest padding with Straps	770194	Retrofit to Size B standard backrest for wedges
25mm Thoracic Backrest padding with Straps	770193	Retrofit to Size C standard backrest for wedges
Small Thoracic Wedges (Singular)	770042	Positional aids to support user while sitting in chair
Pair Medium Thoracic Wedges (Singular)	770041	Positional aids to support user while sitting in chair
Pair Large Thoracic Wedges (Singular)	770040	Positional aids to support user while sitting in chair
Footrest Accessories		
Description	**Part number**	**Comments**
Pair Standard Footrests Raised 60mm	910169	Footrests raised higher than standard
Pair Standard Footrests Raised 90mm	910174	Footrests raised higher than standard
Pair Standard Footrests Raised 120mm	910179	Footrests raised higher than standard
Heavy Duty Footrests	910148	Aluminium footplate harder wearing
540mm-Clip on padded foot tray	910264	For size A width frame
490mm-Clip on padded foot tray	910266	For size B width frame
440mm-Clip on padded foot tray	910263	For size C width frame
Individual clip-on padded foot tray	910269	Make footplate comfortable for user's feet
Dartex footrest covers	880406	Make footplate comfortable for user's feet
Wrap around Dartex Footrest Covers	880432	Make footrest comfortable for user's feet
540mm -Vinyl calf support	920019	For size A width frame
490mm -Vinyl calf support	920020	For size B width frame
440mm -Vinyl calf support	920021	For size C width frame
Armrest Accessories		
Description	**Part number**	**Comments**
Pair Clip on 25mm Padding for Armrests	920056	Not compatible with wide armrests
Pair Clip on 50mm Padding for Armrests	920057	Not compatible with wide armrests
Pair Clip on 75mm Padding for Armrests	920058	Not compatible with wide armrests

Hygiene		
Description	**Part number**	**Comments**
Oval Commode Pan and Lid	910214	N/A
Cup Shaped Splashguard for a Horseshoe Seat	910216	Suitable for all horseshoe seats
Cup Shaped Splashguard for an Aperture seat	910215	Only suitable for a medium opening aperture seat

Table 3-1

4 Assembly Instructions

The sections below will give detailed instructions on how to install and remove the accessories from your chair. These instructions must be adhered to for a successful and safe assembly of the chair. If you are unsure of any processes, always seek advice.

4.1 Attaching the seat

a. Place the frame on a flat surface and apply the brakes.
b. Remove the plastic bag containing the black star handle nuts attached to the side of the frame. Keep these to one side as they are needed to secure the armrests into position.
c. Remove the seat from its packaging and place it onto the frame in the orientation shown below in Fig 4.1-1.
d. Rotate the 4 plastic retention clips over the frame. Ensure the clips are rotated to the 90°position so the seat is secured to the frame correctly, as shown below in Fig. 4.1-2.



Fig. 4.1-1




Fig. 4.1-2

4.2 Attaching the backrest

a. Remove the backrest from its packaging and unscrew the two-barrel nuts from the backrest.
b. Fit the backrest into the rearmost sockets on the frame, with the handles facing backwards as shown below in Fig. 4.2-1 and 4.2-2.
c. Fit the two-barrel nuts onto the threads at the bottom of the backrest and tighten them up, as shown below in Fig. 4.2-3.



Fig. 4.2-1



Fig. 4.2-2



Fig. 4.2-3

4.3 Attaching the armrest

a. Remove the armrests from their packaging and remove the 2 star handle black nuts from the plastic bag that was previously removed from the frame.
b. Insert the armrest legs into the sockets, ensuring the rivet is at the front, as shown below in Fig 4.3-1.
c. Tighten the star handle black nuts to secure the armrests in place, as shown below in Fig 4.3-2.
d. To swing the armrest away, loosen the black star handle shown in Fig.4.3-2, then lift the armrest out of the front socket and fold backwards, as shown below in Fig. 4.3-3.



Fig. 4.3-1




Fig. 4.3-2



Fig. 4.3-3

4.4 Attaching the footrest

a. To secure footrests in the forward position, place them onto the retaining spigots at the front of the shower chair, as shown below in Fig. 4.4-1.
b. Engage the hook at the side of the footrest over and onto the frame tube, as shown below in Fig. 4.4-2.
c. To fold the footrest back, lift them up until the hook clears the frame and then fold them backwards, as shown below in Fig. 4.4-3.



Fig. 4.4-1



Fig. 4.4-2



Fig. 4.4-3

4.5 Attaching accessories

Below are instructions on how to attach all available accessories to the chair in the correct manner. See relevant accessories below:

4.5.1 Buckle or VELCRO ® Brand fastener Waist Strap

a. Remove the strap from the packaging.
b. You'll find a short strap loop in the middle of the strap. Unfasten the square hook and loop patch on one side to pass the shorter strap around the backrest. Secure it in place by reattaching the hook and loop patch, creating a retaining loop, as shown below in Fig. 4.5.1-1.
c. This allows the strap to remain in place when not in use as shown below in Fig. 4.5.1-2.



Fig. 4.5.1-1



Fig. 4.5.1-2

4.5.2 Full Adult Safety Harness

a. Remove the Harness from the packaging.
b. Loosen the seat clips from the frame: you can refer to the seat fitting instructions.
c. Attach the pelvic buckle straps by unthreading the strap from the buckles and looping them around the sides of the frame. Then thread the straps back through the buckle, pulling the loop tight against the bar, as shown below in Fig. 4.5.2-1. and Fig. 4.5.2-2.
d. Secure the seat clips back into their original position.
e. Attach the shoulder buckle straps to the backrest at the top of the vinyl like the pelvic strap described above and ensure the buckle adjustment is at the rear of the backrest shown below in Fig. 4.5.2-3.
f. Attach the anchor strap to the rear bar behind the seat and secure in place with hook and loop fastening shown below in Fig. 4.5.2-4.


Fig. 4.5.2-1


Fig. 4.5.2-2


Fig. 4.5.2-3


Fig. 4.5.2-4

4.5.3 Toe Straps with Velcro ® Brand fastener

a. Remove the Toe Straps from the packaging.
b. Detach the hook and loop fastening on the top and bottom of the strap, enabling the straps to be threaded through the slots on the footrest. Make sure the longer of the two straps is threaded to the outer side of the footrests, as shown below in Fig. 4.5.3-1.
c. Secure the lower of the two straps to the hook and loop fastening underneath the footrests, as shown below in Fig. 4.5.3-2.
d. Place the user's foot on the footplate and secure the straps over the foot with the hook and loop fastening as shown in figs. 4.5.3-3 and 4.5.3-4.


Fig. 4.5.3-1



Fig. 4.5.3-2


Fig. 4.5.3-3


Fig. 4.5.3-4

4.5.4 Heel Strap/ Heel and Toe Strap combine

a. Remove the Heel Strap or combined Heel and Toe Strap from the packaging.
b. Both the Heel Strap and combined Heel and Toe Straps are attached to the footplate in the same way. Please see Fig. 4.5.4-1 and Fig. 4.5.4-2 below for clarification on the difference between these products.
c. Thread each of the straps through the slot on each side of the footplate, as shown below in Fig. 4.5.4-3.
d. Securing the strap underneath the footplates with hook and loop fastening shown below in Fig,4.5.4-4.
e. If you have the Heel and Toe Strap combination, place your foot onto the heel support on the footplate and secure the straps over the foot with the hook and loop fastening as shown above in Figs. 4.5.3-3 and 4.5.3-4.



Fig. 4.5.4-1 – Heel Support



Fig. 4.5.4-2 – Heel & Toe Support



Fig. 4.5.4-3



Fig. 4.5.4-3

4.5.5 Ankle Supports

a. Remove the Supports from the packaging.
b. Thread the strap with varying width webbing through the slots located on either side of the footplate, as shown below in Fig. 4.5.5-1.
c. Secure the larger straps together underneath the footplate with hook and loop fastening, as shown below in Fig. 4.5.5-2.
d. Attach the support by threading the remaining straps to the buckles situated on each side of the support, as shown in below in Fig. 4.5.5-3.



Fig. 4.5.5-1



Fig. 4.5.5-2



Fig. 4.5.5-3

4.5.6 Calf and Heel Support

a. Remove the Calf and Heel support from the packaging.
b. Detach the strap from the hook and loop fastening to allow it to be threaded around the bottom of the frame, as shown below in Fig. 4.5.6-1.
c. Securing the strap in place by re-attaching it to the hook and loop fastening.



Fig. 4.5.6-1

4.5.7 Dartex Footrest Covers

a. Remove the covers from the packaging.
b. Slot the open end of the Dartex Cover around the footplate, as shown below in Fig. 4.5.7-1.



Fig. 4.5.7-1

4.5.8 Wrap around Dartex Footrest Covers

a. Remove the covers from the packaging.
b. Detach the cover from the hook and loop fastening and thread the strap around the footrest, passing it between the heel support shown below in Fig. 4.5.8.-1.
c. Secure to the footrest by re-attaching the hook and loop fastening back together underneath the footplate, shown below in Fig. 4.5.8-2.



Fig. 4.5.8-1



Fig. 4.5.8-2

4.5.9 Pelvic Strap

a. Remove the strap from the packaging.
b. Thread the 2 adjustment straps around the rear crossbar of the backrest, as shown below in Fig. 4.5.9-1.
c. Secure in place by threading the straps back through the buckle shown below in Fig. 4.5.9-2 and Fig. 4.5.9-3.



Fig. 4.5.9-1



Fig. 4.5.9-2



Fig. 4.5.9-3

4.5.10 Vinyl Calf Support

a. Remove the support from the packaging.
b. Release the buckle clips on the support.
c. Wrap the support around the front of the frame and re-attach the buckle clips, as shown below in Fig. 4.5.10-1.



Fig. 4.5.10-1

4.5.11 Clip on Padded Armrests

a. Remove the armrest padding from the packaging.
b. Align the cutout of the Padded Armrests with the legs on the armrests shown below in Fig. 4.5.11-1.
c. Press the padding firmly to the armrest until it clicks into position, as shown below in Fig. 4.5.11-2.



Fig. 4.5.11-1



Fig. 4.5.11-2

4.5.12 Clip-on Padded Foot Trays /Individual Clip-on Padded Trays

a. Remove the tray from the packaging.
b. Both the Padded Foot Tray and Individual Padded Foot Trays fit to the footplate in the same way. Please consult Fig. 4.5.12-1 and Fig. 4.5.12-3 below, for clarification on the difference between these products.
c. Position the Padded Foot Tray or Individual Padded Foot Trays onto the footplates and firmly press down until they click into position, as shown below in Fig. 4.5.12-2 and Fig. 4.5.12-4.



Fig. 4.5.12-1




Fig. 4.5.12-2



Fig. 4.5.12-3



Fig. 4.5.12-4

4.5.13 Detachable Headrests

a. Remove the headrest from the packaging.
b. Position the headrest onto the spigots situated at the top of the backrest, as shown below in Fig. 4.5.13-1.
c. Use the star handle black nuts to securely fasten the headrest in place, as shown below in Fig. 4.5.13-2.



Fig. 4.5.13-1



Fig. 4.5.13-2

4.5.14 Curved Padded Head Support

a. Remove the support from the packaging.
b. Position the Head Support to the hook and loop fastening located on the detachable headrest, as shown below in Fig. 4.5.14-1.
c. Secure in place by pressing the Head Support to the Headrest.



Fig. 4.5.14-1

4.5.15 Head Support Cushion

a. Remove the cushion from the packaging.
b. Unfasten the clips on the straps located on the rear of the padding.
c. Place the Padding to the front of the Headrest shown below in Fig. 4.5.15-1.
d. Secure in place by reattaching the clips around the back of the backrest, as shown below in Fig. 4.5.15-2.



Fig. 4.5.15-1



Fig. 4.5.15-2

4.5.16 Thoracic Back Pad with Strap and Wedges

a. Remove the strap from the packaging.
b. Unfasten the clips on the straps located on the rear of the padding.
c. Place the padding with the hook and loop fastening on the front, to the front of the backrest, as shown below in Fig. 4.5.16-1.
d. Secure in place by reattaching the clips around the back of the backrest, as shown below in Fig. 4.5.16-2.
e. Attach the wedges to the hook and loop fastening located on the backrest padding show below in Fig. 4.5.16-3.
f. Secure in place by pressing the wedges onto the backrest padding.



Fig. 4.5.16-1



Fig. 4.5.16-2



Fig. 4.5.16-3

5 Operating Instructions

This product may be heavy for users and must be lifted with care. Please operate this product with the use of the castors/wheels wherever possible to prevent any injury.

Appropriate training in lifting, handling, and bathing procedures should be undertaken by any person operating this product, along with any other additional aid equipment, for their own and the client's safety and comfort. These instructions are designed to cover the method of using the product.

5.1 Operating the Braked levers

The brakes on the shower chair are individually braked using the levers.

Do not try to push the product when the brakes are activated.

a. To engage the brake, firmly push the brake lever forward until it clicks into place, effectively securing the brake onto the wheel, as shown in Fig. 5.1.-1.
b. Repeat this process for both brakes located on either side of the frame.
c. To release the brakes, pull the brake lever back into the vertical position, ensuring the brake has fully disengaged, as shown in Fig.5.1.-2.



Fig. 5.1-1



Fig. 5.1-2

5.2 Footrest adjustment

a. Gently loosen the star handle black locking nut located at the rear of the hanger shown below in Fig. 5.2-1.
b. Set the footrest to your preferred height and secure it in place using the locking nut shown below in Fig. 5.2-2.



Fig. 5.2-1



Fig. 5.2-2

5.3 Using the Commode Pan (Accessory)

Make sure that the commode pan is empty before transferring the shower chair.
The commode pan has a maximum capacity of 6 litres.

The plastic commode pan is designed to self-locates onto the pan sliders located underneath the seat. This can be done by sliding the pan in and out from rear of the shower chair, aligning it with the sliders on the seat, as shown below in Figures Fig. 5.3-1, Fig. 5.3-2, and Fig. 5.3-3. Before inserting, remember to remove the lid and store in a suitable location.



Fig. 5.3-1



Fig. 5.3-2



Fig. 5.3-3

5.4 How to use the Harnesses

a. After attaching the harness to the chair, please consult section 4 accessory attachment instructions.
b. Customise the strap adjustments on the harness to ensure a comfortable fit for the user.
c. Once the necessary adjustments have been made, secure in place by fastening the buckles together, with the comfort pad underneath shown below in Fig. 5.4-1 and Fig. 5.4-2.



Fig. 5.4-1



Fig. 5.4-2

5.5 How to use and adjust the waist straps

a. Once the straps have been attached to the frame, please consult section 4 accessory attachment instructions.
b. Gently loosen the adjustment straps to enable the buckles to come together while the user is seated in the chair, as shown below in Fig. 5.5-1.
c. After making the required adjustments, firmly secure by fastening the buckles together, as shown below in Fig. 5.5-2.
d. Tighten the adjustment strap to achieve a comfortable fit for the user, as shown below in Fig. 5.5-3.



Fig. 5.5-1



Fig. 5.5-2



Fig. 5.5-3

5.6 How to use and adjust the pelvic straps

a. Once the straps have been attached to the frame, please consult section 4 accessory attachment instructions.
b. Gently loosen the adjustment straps to enable the buckles to come together while the user is seated in the chair.
c. After making the required adjustments, firmly secure by fastening the buckles together, with the comfort pad underneath as shown below in Fig. 5.6-1 and Fig.5.6-2.



Fig. 5.6-1



Fig. 5.6-2

5.7 How to use and adjust the Ankle Supports

a. Once the straps have been attached to the frame, please consult section 4 accessory attachment instructions.
b. Gently place the user's foot on the footrest. Guide the strap through the side buckle, as shown below in Fig. 5.7-1.
c. Adjust it for users' comfort and secure in place using the hook and loop fastening as shown be below in Fig 5.7.2.




Fig. 5.7-1



Fig. 5.7-2

6 Technical Specification

6.1 Shower Chair Frame Dimensions




T60 Specification – millimetre			
Dimensions (Ref to drawing)	**T60A**	**T60B**	**T60C**
A – Maximum Height	980	980	980
B – Maximum Width	640	590	540
C – Internal Armrest Width	470	420	370
D – Internal Height	470	470	470
E – Internal Width	485	435	385
F – Toilet Seat Internal Height Clearance	445	445	445
G – Floor to Seat Height	560	560	560
H – Ground Clearance	206	206	206
I – Maximum Length (Including Footrest)	1050	1050	1050
J – Seat to Armrest Height	180	180	180
K – Seat to Handle Height	450	450	450
L – Seat Depth	400	400	400
M – Centre of Seat to Rear of Wheel	350	350	350
Turning Circle	630	630	100
Castor Size	125	125	125
Wheel Size	510	510	510
Footrest Adjustment	155	155	155

T60 Weight Specifications			
	T60A	**T60B**	**T60C**
Safe Working Load (SWL)	200kg	200kg	200kg
Commode SWL (Capacity)	6L	6L	6L
Product & Packaging Weight	22kg	15kg	15kg
Product Weight	21kg	14kg	14kg


Table 5.1-1

6.2 Shower Chair Aperture Dimensions

Horseshoe Aperture:



Full Ring Aperture:



7 Environmental Conditions

The sections below will give detailed information regarding the environmental conditions the product should maintain throughout its life cycle. Failure to adhere to these conditions may negatively impact the function of the product. If you are unsure of any environmental conditions, always seek advice.

7.1 Operating Environment

The shower chair is intended to be used in both dry and wet environments, with its primary use being in shower rooms. The shower chair is intended for internal use only.

The shower chair is intended to be used within a professional healthcare facility or home healthcare environment. The shower chair is not suitable for any other environment.

The shower chair is not intended to be used in environments where there are rapid changes in the environmental temperature and humidity.

7.1.1 Normal Operating Conditions

+5˚C to +40˚C (41˚F to 104˚F) at a relative humidity between 15% to 90% RH, non-condensing but not requiring a water vapour pressure greater than 50hPa and atmospheric pressure between 700hPa to 1060hPa.

7.2 Storage Conditions

The shower chair is intended for internal storage within normal environmental conditions. The shower chair is intended to be stored in a dry room.
It is not intended to be stored in environments where there are rapid changes in the environmental temperature.
When storing the chair, ensure that the product has been cleaned and dried. For further information, refer to the section 8 cleaning instructions.

7.2.1 Shipping and Storage Conditions

-25˚C to +5˚C (-13°F to 41°F) with any humidity level.
+5˚C to +35˚C (41˚F to 95˚F) at a relative humidity up to 90%.
+35˚C to 70˚C non-condensing at a water vapour pressure up to 50hPa.

8 Daily Checks

Inspection is to be completed prior to each use by the user of the shower chair.



Should any of the components in the table below fail the inspection, DO NOT use the shower chair. Contact your local authorised dealer for service – contact details are on the last page of this manual.

Ensure all component inspections in the table below are completed prior to each use of the shower chair.

Check list before use:

Component	Service/inspection required
Generic	Visual inspection of the external of the shower chair. Significant damage that may affect the shower chair's function and a clear safety hazard is unacceptable.
	Check the product labelling to ensure they are all still legible. This includes the serial number and other important markings. If labels are not legible, then contact your local authorised dealer immediately.
	Check all main nuts and bolts to see if they are loose. If they are not tight or you have concerns, then contact your local authorised dealer immediately.
Seat	Inspect the seat for any damage including tears and excessive wear. If the integrity of the seat is compromised, it should be replaced.
	Ensure that the seat is fixed onto the frame correctly. See section 4 assembly instructions for guidance.
Backrest	Ensure that the vinyl backrest shows no sign of tears or excessive wear. If the integrity of the backrest is compromised, the chair should not be used, and the backrest must be replaced.
Armrest	Inspect the armrest for any damage including tears and excessive wear. If the integrity is compromised it should be replaced.
	Ensure that the armrest is fixed onto the frame correctly. See Section 4 assembly instructions for guidance.
Brakes	Ensure the brakes function when activated.
	Ensure the product can move freely when the brakes are deactivated. (Ensure they aren't rubbing on the wheels).
Castors/Wheels	Examine the 'products' castors/wheels for signs of damage or wear, ensuring they run freely and are not damaged or cracked.
	Check the castor/wheel fixings for hair and dirt etc., as a build-up can restrict the castor/wheel movement.
	Check to see all castors/wheels are flat on the floor and shower chair is stable.
	Ensure the product can rotate freely with ease.

Table 7-1

9 Cleaning



To reduce the risk of cross-contamination it is recommended to clean the shower chair and accessories before use by a different person.

Please follow the recommended cleaning guidelines below on cleaning and disinfecting the shower chair.

9.1 Metal Frame Cleaning

The metal frames of the chairs are powder-coated with a coating containing a silver anti-microbial additive. The coating provides increased scuff and abrasion resistance compared to conventional paint, which will help protect the product from rust etc., while reducing the levels of MRSA bacteria by at least 99.99%.

For mild cleaning the frames can be cleaned using a damp cloth, soap/water, and antibacterial spray. For a deeper clean, a low-pressure steam cleaner can also be used. Do not use a steam cleaner at high pressure, as this could damage the powder coating and labels. Do not use industrial bleaches, abrasive cleaners, or organic solvents.

All cleaning solutions must be thoroughly rinsed off the product at the end of the cleaning process and the product dried using a dry cloth/towel. Always ensure the product is dry before use.

9.2 Fabric Cleaning

The vinyl is a hardwearing, waterproof, and enduring material which provides MRSA bacteria resistance. This material is found on the seat and backrest.

The Dartex material is an alternative material which provides waterproof properties, along with featuring multi-directional stretch for pressure redistribution. The Dartex includes anti-mycotic protection which is impenetrable for bacteria and viruses.

Both materials can be wiped down using a damp cloth to remove any mild dirt, dust, and minor spillages. When a spillage occurs, it is recommended that it be cleaned as quickly as possible to avoid any staining; it is good practice to dab the spillage and not to rub it as this could cause staining. For more persistent stains and dirt, hot water with an antibacterial spray can also be used.

All cleaning solutions must be thoroughly rinsed off the product at the end of the cleaning process and dried using a dry cloth/towel. Always ensure the product is dry before use.

9.3 Disinfecting (if necessary)

Should the shower chair require a more thorough cleaning, the use of the Actichlor™ disinfectant product (which is widely available in tablet form and used throughout the healthcare industry) is recommended.



Follow the manufacturer's safety instructions for the use of this cleaning product before use to ensure safe use for the operator and the user.
Ensure the cloth is damp before the cleaning process.

Application is through a clean damp cloth applied to wipe the product down. Use in the following dilutions to ensure an effective clean:

- Actichlor™ dissolvable chlorine tablets provide a concentration of 1000 ppm of available chlorine (0.1%) per 1 tablet.
- 1 tablet (1.7g formed tablet (x1)) will create a virucidal solution, diluted in 1 litre of water to provide effective means to clean a "dirty" product. This is also ideal for use after an outbreak of the Norovirus/winter vomiting and can be used as a precaution against C.Diff. It is effective against viruses, bacteria, spores, yeasts, and moulds.

- A minimum of 5 minutes contact time with the outer components is recommended to prevent virucidal infections, whilst maintaining the integrity of the product. The product can withstand a longer contact period, however a minimum of 5 minutes is required to provide an effective cleaning regime.
- Blood spills should be dealt with by an increased concentration of the solution – please refer to the instructions on the manufacturer's product labelling.

Dilution chart					
Product used as	**Product condition**	**Concentration (ppm)**	**Dilution qty* (l)**	**Tablets per 1l (0.26gal)**	**Contact time (minutes)**
Bactericidal	Clean	200	5 (1.32gal)	1	1
	Dirty	1000	1 (0.26gal)	1	5
Yeasticidal	Clean	200	5 (1.32gal)	1	1
	Dirty	1000	1 (0.26gal)	1	5
Fungicidal	Clean	2000	1 (0.26gal)	2	15
	Dirty	5000	1 (0.26gal)	5	15
Mycrobactericidal	Clean	1000	1 (0.26gal)	1	15
	Dirty	5000	1 (0.26gal)	5	15
Virucidal	Clean	500	2 (0.53gal)	1	5
	Dirty	1000	1 (0.26gal)	1	5
Sporcidal (C. Diff)	Clean	1000	1 (0.26gal)	1	10
	-	-	-	-	-
Sporcidal	Clean	5000	1 (0.26gal)	5	10
	-	-	-	-	-

- Dilution is made with water. DO NOT dilute within any other medium.
- When diluted in water, one tablet gives 1000ppm of available chlorine.
- The concentration of the solution depends upon whether the object being cleaned is noticeably dirty (indicated in the table by "Product condition".

Table 9.3-1

Handling and storage safety precautions when using this cleaning agent:		
Advice on Safe Handling	**Hygiene Measures**	**Conditions for Safe Storage**
Avoid contact with skin and eyes. Do not breathe dust, fumes, gas, mist, vapours, spray. Use only with adequate ventilation. Wash hands thoroughly after handling. Mixing this product with acid or ammonia releases chlorine gas.	Handle in accordance with good industrial hygiene and safety practice. Remove and wash contaminated clothing before re-use. Wash face, hands, and any exposed skin thoroughly after handling.	Keep out of reach of children. Keep container tightly closed. Store in suitable labelled containers. Storage temperature: 0-25˚C (32-77˚F).
Individual Protective Measures	**Dissolve**	
Hand protection: Gloves	Dissolve in cold water – With no agitation, 1 tablet will take approximately 10 minutes to fully dissolve in the water used.	
The information above has been extracted from the Actichlor™ MSDS (Manufacturers Safety Data Sheet). For a full review of the data please follow the link below: http://www.nhsggc.org.uk/media/236215/msds-actichlor-plus.pdf		

Table 9.3-2

10 Servicing

Regular servicing on the shower chair will help prevent breakdowns and reduce repair costs. It will also improve the quality of the product for the end users.

To reduce the risk of injury, no service is to be carried out on the shower chair while in use.
Service must be completed by an authorised service engineer only.
Do not attempt to service the product yourself; this will void your warranty.

To ensure the safety and continued good function of your shower chair, it is recommended to have an approved service engineer perform a routine service every 12 months, this will ensure that the product meets the required standards. It is important to document the service history of the product in the service log located at the back of this user manual after each service.

When the product is serviced, the service checklist must be completed.
Service Manual Document Number: 995167.

For information regarding spare parts, refer to the spare parts manual.
Spare Parts Manual Document Number: 992167.

Contact your local authorised dealer if you:

- Need more information.
- Have any questions about the use or service of your shower chair.
- Notice any change in the performance.
- Want to report an unexpected occurrence.
- Want to arrange a service.
- Need to ascertain necessary information for replacement parts and components.

The expected product lifetime is **10 years**. This is dependent on usage and compliance with maintenance. Regular service on the product will increase the expected lifetime.

11 Troubleshooting

Should a problem arise with the use of the shower chair, review the Table below.

Find the fault and complete the recommended solution.

If the fault is not listed below or the solution does not correct the problem, contact your local authorised dealer immediately – contact details are provided on the last page of this manual.

Fault	Action
Castors/Wheels do not operate	Ensure the brake has been fully disengaged. Check that the castors/wheels are running freely and are level. Check the castors/wheels are not fouled with debris that would restrict their movement (clean out, test, and return to service). If this does not correct the fault, then contact your local authorised dealer immediately so the product can be checked to ensure proper continued operation.
Brakes do not operate	Ensure that the brake mechanism is fully engaged. Check to see that all castors/wheels are touching the ground. If this does not correct the fault, then contact your local authorised dealer immediately so the product can be checked to ensure proper continued operation.
Footrest adjustment does not hold	If the footrest adjustment does not operate, check the star handle is tight; it may be that the thread has been stripped in the leg rest attachment point. If so, contact your local authorised dealer immediately so the product can be checked to ensure proper continued operation.

Table 11-1

12 Disposal

With efforts to improve the environment and reduce waste, where possible our products have been manufactured with recyclable materials. Below are our guidelines on recyclable materials and being environmentally friendly.

Please observe the local laws on recycling and respect the current laws for disposal within the community the product is being used within. If there is any uncertainty of the below guidelines, contact your local authorities to determine the proper method of disposal of potentially biohazardous parts and accessories.

Fully Recyclables	Consideration when Recycling
All steel framework	Seats
Castors/Wheels	Rubber handlebar covers
Initial packaging of the product (cardboard)	Plastic fixings
Metallic fixings – screws etc.	Harnesses + Belts

Table 12-1

When the shower chair has completed its life cycle and can no longer perform to its intended use safely, the product should be decommissioned by an approved Service Engineer.



The product may be contaminated and must be disinfected before recycling or disposal.
See section on 'Cleaning' for details of how to do this

13 Warranty

It is impossible for all the risks to be eliminated from use of this product, but to reduce risk and improve safe and proper use, the user should always read and understand the user manual before use. Product failure may occur due to lack of maintenance and care, misuse, unauthorised and improper servicing or alterations, improper storage, and environmental use, or through normal use wear and tear. These factors are all beyond the control from the manufacturer. These risks are taken on by the users.

Once registered, the manufacturer will provide a 5-year warranty period for the steelwork and 1-year warranty for all other products against all manufactured defects to the registered customer. The warranty is valid if the product has maintained its intended use and the user manual instructions have been followed. The warranty will not extend to the use of the product when used contrary to the user manual. This guarantee does not affect or in any way limit your statutory rights.

.

1. The liability of the manufacturer under the terms of this guarantee shall be limited to the replacement of the defective part(s) to the sales distributor, dealer, agent, person, or entity which purchased the equipment from the manufacturer. In no event shall the manufacturer incur liability for any consequential or unforeseeable losses.

2. This equipment guarantee shall be void if the equipment is not serviced by Prism Medical UK or its authorised service engineers, in accordance with the manufacturer's recommendations, or if any unauthorised persons carry out work on the equipment.

3. This guarantee does not apply to failure attributable to normal wear and tear, damage by natural forces, user neglect or misuse or deliberate destruction.

4. Do not attempt to service the product yourself, or the warranty is void.

14 Service Record History

Complete this section after each service, repair inspection and/or maintenance.

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	
Date:	Time:
Service Type: Service Inspection ☐ Repair ☐ Other ☐	
Completed By: ..... (Printed name) ..... (Signature) Company:	
Remarks & Actions Taken:	
Product Left in A Safe & Usable Condition: Yes ☐ No ☐ (if no explain in actions above)	

User notes:

Dealer/service contact details:

Manufacturer contact details:

Prism Medical UK
Unit 1, Tir Llwyd Industrial Estate,
St Asaph Avenue, Kinmel Bay,
Conwy, United Kingdom, LL18 5JZ

Telephone Number: 01924 840100

Disclaimer

While every effort has been made to ensure the accuracy of information contained in this user manual, no liability can be accepted by Prism Medical UK for any errors or omissions. Prism Medical UK operates a policy of continuous improvement. Specifications and other data are subject to change without notice.

www.prismmedical.co.uk